

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062118\
 Data File : VU024814.D
 Acq On : 21 Jun 2018 16:29
 Operator : MD/SY
 Sample : J3539-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S39-0183

Quant Time: Jun 22 04:33:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	189430	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	289588	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	269220	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	145081	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	134389	43.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.92%	
35) Dibromofluoromethane	4.89	113	111798	46.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.04%	
50) Toluene-d8	7.57	98	421229	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
62) 4-Bromofluorobenzene	10.31	95	154430	44.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.74%	
Target Compounds						
16) Acetone	2.32	43	10037	6.04	ug/l	Qvalue 100
27) cis-1,2-Dichloroethene	4.23	96	6972	2.55	ug/l	88
44) Trichloroethene	6.19	130	36102	13.08	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062118\
Data File : VU024814.D
Acq On : 21 Jun 2018 16:29
Operator : MD/SY
Sample : J3539-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S39-0183

Quant Time: Jun 22 04:33:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

