

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062118\
 Data File : VU024815.D
 Acq On : 21 Jun 2018 16:53
 Operator : MD/SY
 Sample : J3539-06
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S39-0184

Quant Time: Jun 22 04:38:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	193781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	295365	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	279437	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	146385	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	137648	43.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.02%	
35) Dibromofluoromethane	4.89	113	115423	47.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.18%	
50) Toluene-d8	7.57	98	432839	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
62) 4-Bromofluorobenzene	10.31	95	157526	44.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.74%	
Target Compounds						
16) Acetone	2.32	43	11649	6.85	ug/l	96
27) cis-1,2-Dichloroethene	4.24	96	4330	1.55	ug/l	80
32) 1,1,1-Trichloroethane	4.92	97	5783	1.32	ug/l	94
44) Trichloroethene	6.19	130	2696	0.96	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062118\
Data File : VU024815.D
Acq On : 21 Jun 2018 16:53
Operator : MD/SY
Sample : J3539-06
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S39-0184

Quant Time: Jun 22 04:38:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

