

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062118\
 Data File : VU024816.D
 Acq On : 21 Jun 2018 17:18
 Operator : MD/SY
 Sample : J3561-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S32-0623

Quant Time: Jun 22 04:42:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	197009	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	297839	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	280586	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	153517	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	143067	44.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.96%	
35) Dibromofluoromethane	4.89	113	117718	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
50) Toluene-d8	7.57	98	438929	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
62) 4-Bromofluorobenzene	10.31	95	162012	45.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.52%	
Target Compounds						
4) Vinyl Chloride	1.40	62	2807180	1080.46	ug/l	98
12) 1,1-Dichloroethene	2.29	96	2878	1.28	ug/l	83
21) trans-1,2-Dichloroethene	2.99	96	67024	27.13	ug/l	96
27) cis-1,2-Dichloroethene	4.23	96	1325471	466.64	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062118\
Data File : VU024816.D
Acq On : 21 Jun 2018 17:18
Operator : MD/SY
Sample : J3561-01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S32-0623

Quant Time: Jun 22 04:42:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

