

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062118\
 Data File : VU024817.D
 Acq On : 21 Jun 2018 17:42
 Operator : MD/SY
 Sample : J3561-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S32-0624

Quant Time: Jun 22 04:44:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	199761	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	307027	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	286038	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	153972	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	144557	44.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.66%	
35) Dibromofluoromethane	4.89	113	119864	47.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.08%	
50) Toluene-d8	7.57	98	447565	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
62) 4-Bromofluorobenzene	10.31	95	164197	44.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.98%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	2797367	1061.85	ug/l	98
12) 1,1-Dichloroethene	2.29	96	2801	1.23	ug/l	87
21) trans-1,2-Dichloroethene	2.99	96	67875	27.09	ug/l	94
27) cis-1,2-Dichloroethene	4.23	96	1325861	460.35	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062118\
Data File : VU024817.D
Acq On : 21 Jun 2018 17:42
Operator : MD/SY
Sample : J3561-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S32-0624

Quant Time: Jun 22 04:44:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

