

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062118\
 Data File : VU024819.D
 Acq On : 21 Jun 2018 18:31
 Operator : MD/SY
 Sample : J3552-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MWHR2-10-061418

Quant Time: Jun 22 04:48:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	184880	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	281995	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	267257	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	142259	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	133246	44.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.30%	
35) Dibromofluoromethane	4.89	113	110722	47.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.62%	
50) Toluene-d8	7.57	98	415541	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
62) 4-Bromofluorobenzene	10.31	95	152086	44.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.74%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062118\
Data File : VU024819.D
Acq On : 21 Jun 2018 18:31
Operator : MD/SY
Sample : J3552-01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MWHR2-10-061418

Quant Time: Jun 22 04:48:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

