

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062118\
 Data File : VU024821.D
 Acq On : 21 Jun 2018 19:20
 Operator : MD/SY
 Sample : J3552-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MWHR3-14-061418

Quant Time: Jun 22 04:52:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	188555	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	287796	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	270303	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	141784	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	135896	44.15	ug/l	0.00
Spiked Amount			Recovery	=	88.30%	
35) Dibromofluoromethane	4.89	113	110479	46.25	ug/l	0.00
Spiked Amount			Recovery	=	92.50%	
50) Toluene-d8	7.57	98	416374	47.79	ug/l	0.00
Spiked Amount			Recovery	=	95.58%	
62) 4-Bromofluorobenzene	10.31	95	152170	43.99	ug/l	0.00
Spiked Amount			Recovery	=	87.98%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.32	43	7035	4.25	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062118\
Data File : VU024821.D
Acq On : 21 Jun 2018 19:20
Operator : MD/SY
Sample : J3552-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MWHR3-14-061418

Quant Time: Jun 22 04:52:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

