

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062118\
 Data File : VU024806.D
 Acq On : 21 Jun 2018 13:15
 Operator : MD/SY
 Sample : J3556-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SRI3-TB1

Quant Time: Jun 22 04:01:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	201038	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	305368	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	289113	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	157136	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	142594	43.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.90%	
35) Dibromofluoromethane	4.89	113	120233	47.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.88%	
50) Toluene-d8	7.57	98	447160	48.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.74%	
62) 4-Bromofluorobenzene	10.31	95	165863	45.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.38%	


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062118\
Data File : VU024806.D
Acq On : 21 Jun 2018 13:15
Operator : MD/SY
Sample : J3556-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SRI3-TB1

Quant Time: Jun 22 04:01:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 4.99 min Scan# 1368

Delta R.T. -0.00 min

Lab File: VU024806.D

Acq: 21 Jun 2018 13:15

Instrument :

MSVOA_U

ClientSampleId :


SRI3-TB1

Tot Ion:168 Resp: 201038

Ion Ratio Lower Upper

168 100

99 52.2 43.2 64.8



Abundance Ion 167.90 (167.60 to 168.60): VU024806.D (-1275) (-)

Ion 98.90 (98.60 to 99.60): VU024806.D (-1275) (-)

4.99

80000



60000

40000

20000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 43.45 ug/l

RT: 5.31 min Scan# 1469

Delta R.T. -0.00 min

Lab File: VU024806.D


Acq: 21 Jun 2018 13:15

Tgt Ion: 65 Resp: 142594

Ion Ratio Lower Upper

65 100

67 52.7 0.0 107.0



Abundance Ion 64.90 (64.60 to 65.60): VU024806.D (-1376) (-)

Ion 66.90 (66.60 to 67.60): VU024806.D (-1376) (-)

5.31

80000


60000


40000

20000

0

Time-->





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 5.89 min Scan# 1649

Delta R.T. -0.00 min

Lab File: VU024806.D

Acq: 21 Jun 2018 13:15

Instrument :

MSVOA_U

ClientSampleId :

SRI3-TB1


Tot Ion:114 Resp: 305368

Ion Ratio Lower Upper

114 100

63 20.5 0.0 45.4




88 15.9 0.0 31.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VU0

Ion 87.90 (87.60 to 88.60): VU0



#35

Dibromofluoromethane

Concen: 47.44 ug/l

RT: 4.89 min Scan# 1337

Delta R.T. 0.00 min

Lab File: VU024806.D

Acq: 21 Jun 2018 13:15


Tgt Ion:113 Resp: 120233

Ion Ratio Lower Upper

113 100

111 101.4 82.2 123.4



192 20.8 16.2 24.4



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 48.37 ug/l
RT: 7.57 min Scan# 2171
Delta R.T. -0.00 min
Lab File: VU024806.D
Acq: 21 Jun 2018 13:15




Instrument :
MSVOA_U
ClientSampleId :
SRI3-TB1

Tot Ion: 98 Resp: 447160
Ion Ratio Lower Upper
98 100
100 63.8 51.1 76.7



#62
4-Bromofluorobenzene
Concen: 45.19 ug/l
RT: 10.31 min Scan# 3023
Delta R.T. -0.00 min
Lab File: VU024806.D
Acq: 21 Jun 2018 13:15

Tgt Ion: 95 Resp: 165863
Ion Ratio Lower Upper
95 100
174 85.6 0.0 165.8
176 83.7 0.0 159.4









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 9.09 min Scan# 2644
Delta R.T. -0.00 min
Lab File: VU024806.D
Acq: 21 Jun 2018 13:15

Instrument :
MSVOA_U
ClientSampleId :
SRI3-TB1

Tot Ion	Ratio	Lower	Upper
117	100		
82	56.0	44.3	66.5
119	31.8	25.4	38.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 11.49 min Scan# 3389
Delta R.T. -0.00 min
Lab File: VU024806.D
Acq: 21 Jun 2018 13:15

Tgt Ion	Ratio	Lower	Upper
152	100		
115	57.0	43.0	129.0
150	159.0	0.0	354.0

