

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062122\
 Data File : VU049325.D
 Acq On : 21 Jun 2022 14:20
 Operator : SY/MD
 Sample : N3400-21
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AF6

Quant Time: Jun 22 01:48:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jun 22 01:47:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	131303	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	126200	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	61694	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	35185	3.067	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	61.400%	
7) Chloroethane-d5	1.916	69	29877	3.846	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	77.000%	
11) 1,1-Dichloroethene-d2	2.568	65	13481	2.921	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	58.400%#	
20) 2-Butanone-d5	4.636	46	158656	51.791	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	103.580%	
24) Chloroform-d	5.067	84	79310	4.406	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.200%	
26) 1,2-Dichloroethane-d4	5.703	65	45466	4.604	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	92.000%	
32) Benzene-d6	5.729	84	146461	3.852	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	77.000%	
36) 1,2-Dichloropropane-d6	6.690	67	53301	4.131	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	82.600%	
41) Toluene-d8	7.899	98	120680	3.525	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	70.600%	
43) trans-1,3-Dichloroprop...	8.179	79	18974	3.310	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	66.200%	
46) 2-Hexanone-d5	8.636	63	102542	44.343	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	88.680%	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	45488	4.403	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	88.000%	
66) 1,2-Dichlorobenzene-d4	12.195	152	51594	4.358	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	87.200%	
Target Compounds						
					Qvalue	
14) Carbon disulfide	2.790	76	3685	0.151	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062122\
Data File : VU049325.D
Acq On : 21 Jun 2022 14:20
Operator : SY/MD
Sample : N3400-21
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AF6

Quant Time: Jun 22 01:48:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jun 22 01:47:02 2022
Response via : Initial Calibration

