

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062123\
 Data File : VU054634.D
 Acq On : 21 Jun 2023 11:28
 Operator : MD/SY
 Sample : 03326-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCA32

Quant Time: Jun 22 02:43:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 22 02:42:14 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	310289	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	295997	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	144429	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	118202	5.986	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	119.800%	
7) Chloroethane-d5	1.914	69	91925	4.972	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	99.400%	
11) 1,1-Dichloroethene-d2	2.563	65	44984	5.100	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	102.000%	
20) 2-Butanone-d5	4.644	46	246061	44.319	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	88.640%	
24) Chloroform-d	5.058	84	176778	4.283	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	85.600%	
26) 1,2-Dichloroethane-d4	5.698	65	87629	4.355	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.200%	
32) Benzene-d6	5.724	84	357012	4.339	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.800%	
36) 1,2-Dichloropropane-d6	6.685	67	117459	4.275	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	85.400%	
41) Toluene-d8	7.894	98	302274	4.139	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.800%	
43) trans-1,3-Dichloroprop...	8.177	79	43540	4.556	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	91.200%	
46) 2-Hexanone-d5	8.631	63	162949	40.646	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	81.300%	
56) 1,1,2,2-Tetrachloroeth...	10.753	84	84649	4.131	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	82.600%	
66) 1,2-Dichlorobenzene-d4	12.187	152	107831	4.426	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	88.600%	
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.576	96	36887	1.639	ug/L	# 15
18) trans-1,2-Dichloroethene	3.354	96	3006	0.129	ug/L	92
19) 1,1-Dichloroethane	3.869	63	15789	0.328	ug/L	99
22) cis-1,2-Dichloroethene	4.663	96	129535	4.895	ug/L	98
34) Trichloroethene	6.541	95	14549118	517.203	ug/L	93
45) 1,1,2-Trichloroethane	8.396	97	2814	0.154	ug/L	92
51) Chlorobenzene	9.441	112	10966	0.161	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062123\
Data File : VU054634.D
Acq On : 21 Jun 2023 11:28
Operator : MD/SY
Sample : 03326-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCA32

Quant Time: Jun 22 02:43:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 22 02:42:14 2023
Response via : Initial Calibration

