


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062123\
Data File : VU054630.D
Acq On : 21 Jun 2023 09:00
Operator : MD/SY
Sample : BFB023
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB023

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Title : TRACE VOA SFAM1.0
Last Update : Fri Jun 23 04:31:37 2023



AutoFind: Scans 2904, 2905, 2906; Background Corrected with Scan 2889

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	43568	PASS
75	95	30	80	52.4	114861	PASS
95	95	100	100	100.0	219008	PASS
96	95	5	9	6.2	13537	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	76.8	168107	PASS
175	174	5	9	8.0	13468	PASS
176	174	95	101	95.3	160256	PASS
177	176	5	9	6.4	10333	PASS

m/z	Abundance
35.80	2073.0
36.80	144.0
37.80	795.0
39.90	205.0
43.80	910.0

Instrument :
MSVOA_U
ClientSampleId :
BFB023

m/z	Abundance
35.90	4134.0
37.00	10589.0
38.00	10651.0
39.00	5052.0
39.90	464.0
43.00	154.0
43.90	2072.0
45.00	2102.0
46.00	206.0
46.90	2891.0
47.80	1244.0
48.90	9866.0
50.00	45296.0
51.00	13689.0
52.00	605.0
53.00	343.0
53.90	133.0
54.90	548.0
55.90	3354.0
57.00	5509.0
58.00	257.0
60.00	2654.0
61.00	11966.0
62.00	11181.0
63.00	9825.0
63.90	1364.0
65.00	2451.0
66.00	314.0
67.00	822.0
68.00	22256.0
69.00	24296.0
70.00	1712.0
72.00	1227.0
72.90	10554.0
74.00	38448.0
75.00	116328.0
76.00	9402.0
76.90	1267.0
78.00	1281.0
78.90	8017.0
79.90	2062.0
80.90	7408.0
81.90	1678.0
82.80	196.0
85.80	354.0
86.90	10631.0
87.90	9382.0
90.90	901.0
91.90	6276.0
92.90	9977.0
94.00	25872.0
95.00	221952.0
96.00	14471.0
97.00	516.0
102.80	112.0
103.80	1088.0
104.90	372.0
105.80	1017.0
106.90	238.0
109.90	215.0
110.80	255.0
111.80	185.0
112.70	227.0
114.90	269.0
115.80	928.0
116.90	1412.0
117.80	957.0
118.90	1373.0
127.80	897.0
128.80	441.0
129.90	820.0
130.70	289.0
134.80	403.0
135.80	116.0
136.80	430.0
140.00	155.0
140.90	2624.0
142.80	3044.0
143.80	133.0
144.90	584.0
145.70	340.0
147.80	487.0
148.80	151.0
149.90	135.0
152.70	213.0
153.60	102.0
154.90	600.0
155.80	199.0
156.70	284.0
157.10	320.0
157.90	101.0
158.90	234.0

Instrument :
MSVOA_U
ClientSampleId :
BFB023

160.80	129.0
170.50	246.0
171.00	324.0
171.90	3373.0
173.90	168320.0
174.90	13655.0
175.90	159296.0
176.90	10214.0
177.80	354.0

Instrument :
MSVOA_U
ClientSampleId :
BFB023

m/z	Abundance
35.90	4158.0
37.00	10810.0
38.00	10532.0
38.90	5111.0
39.90	485.0
43.10	103.0
43.90	2137.0
45.00	2175.0
45.80	224.0
46.90	3066.0
47.80	1188.0
48.90	10008.0
50.00	44656.0
51.00	13662.0
52.00	523.0
53.00	258.0
53.90	114.0
54.90	599.0
55.90	3219.0
57.00	5275.0
57.90	287.0
60.00	2567.0
61.00	11811.0
62.00	11209.0
63.00	9308.0
63.90	1409.0
65.00	2639.0
65.90	299.0
67.00	791.0
68.00	22992.0
69.00	24616.0
69.90	1760.0
72.00	1237.0
73.00	10985.0
74.00	38512.0
75.00	118600.0
76.00	9626.0
77.00	1453.0
78.00	1105.0
78.90	8231.0
79.90	2143.0
80.90	7702.0
81.90	1769.0
82.90	209.0
85.80	342.0
86.90	10833.0
87.90	9139.0
90.90	1040.0
91.90	6449.0
93.00	10266.0
94.00	25840.0
95.00	225728.0
96.00	13803.0
97.00	506.0
102.70	121.0
103.80	1098.0
104.90	307.0
105.80	1020.0
106.90	241.0
109.90	269.0
110.80	250.0
111.70	156.0
112.70	248.0
115.00	277.0
115.80	914.0
116.90	1537.0
117.80	1112.0
118.90	1314.0
123.90	124.0
127.80	879.0
128.80	440.0
129.70	893.0
130.80	277.0
134.90	352.0
136.70	388.0
139.80	163.0
140.90	2724.0
142.80	3126.0
143.80	135.0
145.00	461.0
145.70	388.0
147.00	107.0
147.70	428.0
148.80	189.0
149.80	188.0
152.00	108.0
152.70	213.0
154.90	651.0
155.80	167.0
157.00	309.0
157.80	111.0
158.90	283.0

Instrument :
MSVOA_U
ClientSampleId :
BFB023

170.00	109.0
171.00	304.0
171.90	3619.0
173.90	173248.0
174.90	13900.0
175.90	164864.0
176.90	10735.0
177.80	332.0

Instrument :
MSVOA_U
ClientSampleId :
BFB023

m/z	Abundance
35.90	3958.0
37.00	10053.0
37.90	9646.0
39.00	4790.0
40.00	502.0
43.90	1992.0
45.00	1961.0
45.80	300.0
46.90	2771.0
47.80	1045.0
48.90	8834.0
50.00	40752.0
51.00	12756.0
52.00	434.0
52.90	212.0
54.90	739.0
55.90	3086.0
57.00	4734.0
57.90	317.0
60.00	2228.0
61.00	10630.0
62.00	10149.0
63.00	8182.0
63.90	1254.0
65.00	2743.0
66.00	266.0
66.90	725.0
68.00	21632.0
69.00	22232.0
69.90	1728.0
72.00	1207.0
73.00	10460.0
74.00	35584.0
75.00	109656.0
76.00	8973.0
77.00	1494.0
78.00	854.0
78.90	7487.0
79.90	1958.0
80.90	7417.0
81.90	1600.0
82.90	169.0
85.80	304.0
86.90	9806.0
87.90	8094.0
90.90	1014.0
91.90	6028.0
93.00	9620.0
94.00	23608.0
95.00	209344.0
96.00	12337.0
97.00	356.0
102.80	118.0
103.80	1018.0
104.80	252.0
105.80	1003.0
106.90	206.0
109.90	247.0
110.80	222.0
112.00	130.0
112.80	224.0
115.00	236.0
115.90	818.0
116.90	1481.0
117.80	1027.0
118.80	1245.0
127.80	807.0
128.80	395.0
129.80	830.0
130.80	261.0
134.90	314.0
136.80	309.0
139.70	186.0
140.90	2496.0
142.80	2815.0
144.10	136.0
144.70	455.0
145.10	375.0
145.70	392.0
147.00	113.0
147.70	348.0
148.70	160.0
149.60	283.0
152.70	146.0
154.90	584.0
156.00	110.0
156.90	301.0
157.60	145.0
158.90	279.0
170.00	166.0
170.90	276.0
171.90	3335.0

Instrument :
MSVOA_U
ClientSampleId :
BFB023

173.90	162752.0
174.90	12849.0
175.90	156608.0
176.90	10050.0
177.80	301.0

Instrument :
MSVOA_U
ClientSampleId :
BFB023