

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062219\
 Data File : VU032936.D
 Acq On : 21 Jun 2019 11:48
 Operator : JC/SP
 Sample : VU0622WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK36

Quant Time: Jun 22 01:17:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 22 00:32:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	194722	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	193853	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	89895	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	62226	46.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.52%
7) Chloroethane-d5	1.68	69	53179	50.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.22%
11) 1,1-Dichloroethene-d2	2.26	63	96762	36.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.18%
21) 2-Butanone-d5	4.17	46	117020	105.65	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.65%
24) Chloroform-d	4.64	84	129194	49.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.88%
26) 1,2-Dichloroethane-d4	5.30	65	92864	53.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.78%
32) Benzene-d6	5.33	84	254125	50.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.58%
36) 1,2-Dichloropropane-d6	6.32	67	82541	51.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.54%
41) Toluene-d8	7.56	98	232450	49.12	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.24%
43) trans-1,3-Dichloropropene-	7.84	79	40571	48.45	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.90%
47) 2-Hexanone-d5	8.31	63	72447	96.38	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.38%
57) 1,1,2,2-Tetrachloroethane-	10.42	84	121717	48.91	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.82%
64) 1,2-Dichlorobenzene-d4	11.86	152	93958	54.01	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.02%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062219\
Data File : VU032936.D
Acq On : 21 Jun 2019 11:48
Operator : JC/SP
Sample : VU0622WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK36

Quant Time: Jun 22 01:17:43 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 22 00:32:21 2019
Response via : Initial Calibration

