

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062219\
 Data File : VU032948.D
 Acq On : 21 Jun 2019 16:47
 Operator : JC/SP
 Sample : K3469-15
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T11-MW-2-8.0-062019

Quant Time: Jun 22 01:55:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 22 00:32:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	193744	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	188192	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	89393	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	58660	44.30	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.60%
7) Chloroethane-d5	1.68	69	45516	43.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.20%
11) 1,1-Dichloroethene-d2	2.26	63	82685	31.00	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.00%
21) 2-Butanone-d5	4.17	46	124762	113.21	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	113.21%
24) Chloroform-d	4.64	84	117319	45.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.16%
26) 1,2-Dichloroethane-d4	5.30	65	82504	48.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.24%
32) Benzene-d6	5.33	84	221159	45.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.16%
36) 1,2-Dichloropropane-d6	6.33	67	74702	47.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.60%
41) Toluene-d8	7.56	98	195294	42.51	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.02%
43) trans-1,3-Dichloropropene-	7.85	79	36029	44.32	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.64%
47) 2-Hexanone-d5	8.31	63	74031	101.45	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.45%
57) 1,1,2,2-Tetrachloroethane-	10.42	84	116405	48.18	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.36%
64) 1,2-Dichlorobenzene-d4	11.86	152	84769	49.00	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.00%

Target Compounds

					Ovalue
5) Vinyl chloride	1.40	62	888741	505.151 ug/L	99
13) Acetone	2.32	43	1486	0.976 ug/L	77
14) Carbon disulfide	2.47	76	2357	0.611 ug/L #	88
17) trans-1,2-Dichloroethene	2.97	96	6190	4.679 ug/L	97
20) cis-1,2-Dichloroethene	4.22	96	1304821	905.289 ug/L	93
34) Trichloroethene	6.18	95	3058	2.111 ug/L	94
42) Toluene	7.64	91	2801	0.478 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062219\
Data File : VU032948.D
Acq On : 21 Jun 2019 16:47
Operator : JC/SP
Sample : K3469-15
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T11-MW-2-8.0-062019

Quant Time: Jun 22 01:55:28 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 22 00:32:21 2019
Response via : Initial Calibration

