

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062219\
 Data File : VU032982.D
 Acq On : 22 Jun 2019 06:35
 Operator : JC/SP
 Sample : K3464-20
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETF59

Quant Time: Jun 24 02:27:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jun 24 02:01:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	197497	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	190950	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	92138	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	48689	36.07	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	72.14%
7) Chloroethane-d5	1.68	69	45344	42.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.24%
11) 1,1-Dichloroethene-d2	2.26	63	81772	30.07	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.14%
21) 2-Butanone-d5	4.17	46	125759	111.94	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.94%
24) Chloroform-d	4.64	84	121520	46.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.62%
26) 1,2-Dichloroethane-d4	5.30	65	87254	49.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.84%
32) Benzene-d6	5.33	84	226469	45.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.00%
36) 1,2-Dichloropropane-d6	6.32	67	75833	47.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.64%
41) Toluene-d8	7.56	98	205248	44.03	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.06%
43) trans-1,3-Dichloropropene-	7.84	79	33708	40.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.74%
47) 2-Hexanone-d5	8.31	63	72455	97.86	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.86%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	116645	47.58	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.16%
64) 1,2-Dichlorobenzene-d4	11.86	152	83413	46.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.56%

Target Compounds

					Ovalue
5) Vinyl chloride	1.40	62	5872	3.274 ug/L	86
13) Acetone	2.32	43	1959	1.263 ug/L	83
17) trans-1,2-Dichloroethene	2.98	96	2934	2.175 ug/L	98
20) cis-1,2-Dichloroethene	4.22	96	501115	341.068 ug/L	96
34) Trichloroethene	6.18	95	1026566	698.472 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062219\
Data File : VU032982.D
Acq On : 22 Jun 2019 06:35
Operator : JC/SP
Sample : K3464-20
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETF59

Quant Time: Jun 24 02:27:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Jun 24 02:01:10 2019
Response via : Initial Calibration

