


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062220\
 Data File : VU039000.D
 Acq On : 22 Jun 2020 10:15
 Operator : SY/MD
 Sample : BFB18
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB18

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062220WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Mon Jun 22 13:07:40 2020



AutoFind: Scans 2895, 2896, 2897; Background Corrected with Scan 2881

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.9	30466	PASS
75	95	30	80	52.6	76640	PASS
95	95	100	100	100.0	145813	PASS
96	95	5	9	6.1	8957	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	75.4	109962	PASS
175	174	5	9	8.0	8782	PASS
176	174	95	101	96.2	105794	PASS
177	176	5	9	6.3	6629	PASS

m/z	Abundance
36.00	1190.0
37.00	7224.0
38.00	7942.0
39.00	3053.0
40.10	503.0
43.10	192.0
44.00	1513.0
45.00	1253.0
46.10	204.0
47.00	2185.0
48.00	866.0
49.00	6779.0
50.00	30480.0
51.10	9203.0
52.00	476.0
53.10	226.0
55.00	477.0
56.00	2773.0
57.00	4344.0
57.90	124.0
60.10	1459.0
61.00	7961.0
62.00	7830.0
63.00	6845.0
64.10	1165.0
65.00	2008.0
67.00	468.0
68.00	16480.0
69.00	15640.0
70.00	1234.0
72.00	691.0
73.10	7359.0
74.00	27360.0
75.00	76472.0
76.10	6600.0
77.00	955.0
77.80	770.0
78.90	5589.0
79.90	1827.0
80.90	5762.0
81.90	1754.0
83.10	183.0
86.00	203.0
87.00	7003.0
88.00	5980.0
90.90	481.0
92.00	4439.0
93.00	6977.0
94.10	18184.0
95.10	145600.0
96.10	9023.0
97.00	245.0
103.90	950.0
105.00	286.0
105.90	597.0
107.10	168.0
110.90	139.0
111.90	126.0
112.80	141.0
114.90	142.0
115.90	497.0
116.90	1227.0
118.00	642.0
118.90	1159.0
123.90	171.0
127.90	508.0
128.90	244.0
129.90	645.0
130.90	279.0
135.00	388.0
137.00	322.0
141.00	1546.0
142.10	183.0
143.00	2129.0
143.90	159.0
145.00	737.0
146.00	336.0
148.10	289.0
148.90	154.0
149.90	115.0
155.00	408.0
156.90	212.0
158.90	245.0
160.90	186.0
171.10	337.0
172.10	2086.0
174.00	108736.0
175.10	8554.0
176.00	104128.0
177.00	6442.0
178.00	167.0

Instrument :
MSVOA_U
ClientSampleId :
BFB18

m/z	Abundance
35.90	1230.0
37.00	7371.0
38.00	7820.0
39.00	3018.0
39.90	469.0
43.10	221.0
43.90	1743.0
45.00	1252.0
46.00	269.0
47.00	2313.0
48.00	1042.0
49.00	6697.0
50.00	31392.0
51.10	9062.0
52.00	411.0
53.10	205.0
55.00	603.0
56.00	2761.0
57.00	4269.0
58.00	121.0
60.10	1379.0
61.00	7887.0
62.00	7846.0
63.00	6822.0
64.10	1136.0
65.00	1914.0
66.90	467.0
68.00	16528.0
69.00	16178.0
70.00	1221.0
72.00	647.0
73.10	7762.0
74.00	28168.0
75.00	79024.0
76.10	6731.0
77.00	873.0
77.80	699.0
78.90	5684.0
79.90	1738.0
81.00	6005.0
81.90	1828.0
83.10	190.0
86.10	270.0
87.00	6732.0
88.00	6146.0
90.90	598.0
92.00	4416.0
93.00	7135.0
94.10	18704.0
95.10	150144.0
96.10	9231.0
97.00	265.0
103.90	952.0
105.00	268.0
105.90	641.0
107.10	146.0
110.90	187.0
111.90	145.0
112.80	152.0
114.90	141.0
116.00	512.0
116.90	1196.0
117.90	578.0
118.90	1318.0
123.90	137.0
127.90	537.0
128.90	262.0
129.90	617.0
130.90	259.0
135.00	376.0
137.00	327.0
141.00	1540.0
142.00	227.0
143.00	2372.0
143.90	209.0
145.00	663.0
145.90	375.0
148.00	328.0
149.00	113.0
149.90	147.0
154.90	433.0
156.90	229.0
158.90	147.0
160.90	204.0
171.00	342.0
172.10	2039.0
174.00	113384.0
175.10	9015.0
176.00	109728.0
177.00	6761.0
178.00	178.0

Instrument :
MSVOA_U
ClientSampleId :
BFB18

m/z	Abundance
35.90	1194.0
37.00	6992.0
38.00	6874.0
39.00	2857.0
39.80	465.0
43.90	1798.0
45.00	1160.0
46.00	244.0
47.00	2173.0
48.00	1041.0
49.00	6262.0
50.00	29528.0
51.00	8392.0
52.20	359.0
53.10	156.0
55.10	656.0
56.00	2422.0
57.00	4121.0
60.10	1200.0
61.10	7136.0
62.00	7178.0
63.00	6168.0
64.10	1008.0
65.00	1606.0
66.20	138.0
66.90	443.0
68.00	15103.0
69.00	14887.0
70.00	1137.0
72.00	658.0
73.00	7320.0
74.00	26448.0
75.00	74424.0
76.10	6310.0
77.00	690.0
77.80	565.0
78.90	5294.0
80.00	1534.0
80.90	5658.0
81.90	1644.0
83.00	164.0
86.10	314.0
87.00	5887.0
88.00	5587.0
90.90	643.0
92.10	4109.0
93.00	6782.0
94.10	17280.0
95.10	141696.0
96.10	8619.0
97.00	301.0
103.90	805.0
105.00	238.0
106.00	683.0
107.20	109.0
110.00	102.0
111.00	208.0
111.90	119.0
112.80	146.0
114.90	118.0
116.00	537.0
116.90	1119.0
117.90	556.0
118.90	1206.0
128.00	512.0
129.00	218.0
129.90	528.0
130.90	223.0
135.00	350.0
137.00	258.0
141.00	1380.0
142.00	273.0
143.00	2185.0
143.90	210.0
145.00	555.0
145.90	347.0
148.00	331.0
149.90	158.0
154.90	378.0
156.90	214.0
159.00	105.0
160.90	232.0
171.00	289.0
172.10	1912.0
174.00	107768.0
175.10	8779.0
176.00	103528.0
177.00	6685.0
178.00	135.0

Instrument :
MSVOA_U
ClientSampleId :
BFB18

Scan 2881 (10.603 min): VU039000.D
BFB18

m/z	Abundance
39.90	253.0
44.00	914.0

Instrument :
MSVOA_U
ClientSampleId :
BFB18