

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
 Data File : VU049342.D
 Acq On : 22 Jun 2022 11:27
 Operator : SY/MD
 Sample : VU0622WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK146

Quant Time: Jun 23 02:04:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 23 02:04:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	242903	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	228230	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.813	152	104998	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.598	65	60483	40.419	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	80.840%	
7) Chloroethane-d5	1.909	69	47677	43.393	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	86.780%	
11) 1,1-Dichloroethene-d2	2.565	63	90889	32.189	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	64.380%	
21) 2-Butanone-d5	4.623	46	202013	123.131	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	123.130%	
24) Chloroform-d	5.064	84	153741	48.161	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.320%	
26) 1,2-Dichloroethane-d4	5.700	65	102975	49.792	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.580%	
32) Benzene-d6	5.726	84	285643	50.473	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.940%	
36) 1,2-Dichloropropane-d6	6.691	67	106247	55.845	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	111.700%	
41) Toluene-d8	7.896	98	250099	50.109	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	100.220%	
43) trans-1,3-Dichloroprop...	8.179	79	45977	51.368	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	102.740%	
47) 2-Hexanone-d5	8.633	63	124689	125.223	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	125.220%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	165231	51.998	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	104.000%	
66) 1,2-Dichlorobenzene-d4	12.192	152	98715	54.034	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	108.060%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
Data File : VU049342.D
Acq On : 22 Jun 2022 11:27
Operator : SY/MD
Sample : VU0622WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK146

Quant Time: Jun 23 02:04:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 23 02:04:00 2022
Response via : Initial Calibration

