

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
 Data File : VU049343.D
 Acq On : 22 Jun 2022 11:51
 Operator : SY/MD
 Sample : N3394-13
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXEY8

Quant Time: Jun 23 02:04:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 23 02:04:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	241495	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	222513	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	99055	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	59974	40.313	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.620%
7) Chloroethane-d5	1.906	69	50632	46.352	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.700%
11) 1,1-Dichloroethene-d2	2.565	63	86672	30.874	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.740%
21) 2-Butanone-d5	4.617	46	188495	115.561	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	115.560%
24) Chloroform-d	5.063	84	147579	46.501	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.000%
26) 1,2-Dichloroethane-d4	5.700	65	101066	49.153	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.300%
32) Benzene-d6	5.726	84	275673	49.963	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.920%
36) 1,2-Dichloropropane-d6	6.690	67	103781	55.951	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	111.900%
41) Toluene-d8	7.896	98	239714	49.263	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.520%
43) trans-1,3-Dichloroprop...	8.179	79	44142	50.585	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.180%
47) 2-Hexanone-d5	8.629	63	115031	118.492	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	118.490%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	162666	52.506	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.020%
66) 1,2-Dichlorobenzene-d4	12.192	152	91662	53.184	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
Data File : VU049343.D
Acq On : 22 Jun 2022 11:51
Operator : SY/MD
Sample : N3394-13
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXEY8

Quant Time: Jun 23 02:04:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 23 02:04:00 2022
Response via : Initial Calibration

