

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
 Data File : VU049346.D
 Acq On : 22 Jun 2022 13:01
 Operator : SY/MD
 Sample : N3393-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Jun 23 02:05:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 23 02:04:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	234085	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	215828	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	98567	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	57327	39.753	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	79.500%	
7) Chloroethane-d5	1.903	69	48331	45.646	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	91.300%	
11) 1,1-Dichloroethene-d2	2.565	63	82347	30.262	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	60.520%	
21) 2-Butanone-d5	4.620	46	186122	117.718	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	117.720%	
24) Chloroform-d	5.063	84	144093	46.839	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	93.680%	
26) 1,2-Dichloroethane-d4	5.700	65	97605	48.973	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.940%	
32) Benzene-d6	5.726	84	266053	49.713	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.420%	
36) 1,2-Dichloropropane-d6	6.690	67	101117	56.203	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	112.400%	
41) Toluene-d8	7.896	98	226319	47.950	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	95.900%	
43) trans-1,3-Dichloroprop...	8.179	79	41160	48.629	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	97.260%	
47) 2-Hexanone-d5	8.632	63	115378	122.531	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	122.530%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	155181	51.642	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	103.280%	
66) 1,2-Dichlorobenzene-d4	12.192	152	87922	51.266	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	102.540%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
Data File : VU049346.D
Acq On : 22 Jun 2022 13:01
Operator : SY/MD
Sample : N3393-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Jun 23 02:05:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 23 02:04:00 2022
Response via : Initial Calibration

