

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
 Data File : VU049349.D
 Acq On : 22 Jun 2022 14:12
 Operator : SY/MD
 Sample : N3392-03 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXEX8

Quant Time: Jun 23 02:06:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 23 02:04:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	235659	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	223271	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	100539	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	51849	35.714	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	71.420%	
7) Chloroethane-d5	1.896	69	45119	42.328	ug/L	-0.01
Spiked Amount	50.000	Range 70 - 130	Recovery	=	84.660%	
11) 1,1-Dichloroethene-d2	2.562	63	75529	27.571	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	55.140%#	
21) 2-Butanone-d5	4.616	46	173131	108.771	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	108.770%	
24) Chloroform-d	5.063	84	135074	43.614	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.220%	
26) 1,2-Dichloroethane-d4	5.700	65	91898	45.801	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.600%	
32) Benzene-d6	5.726	84	246905	44.597	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	89.200%	
36) 1,2-Dichloropropane-d6	6.690	67	93750	50.371	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	100.740%	
41) Toluene-d8	7.896	98	214381	43.907	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	87.820%	
43) trans-1,3-Dichloroprop...	8.179	79	41754	47.686	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	95.380%	
47) 2-Hexanone-d5	8.632	63	113977	117.008	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	117.010%	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	156420	50.319	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	100.640%	
66) 1,2-Dichlorobenzene-d4	12.192	152	85998	49.161	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	98.320%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
Data File : VU049349.D
Acq On : 22 Jun 2022 14:12
Operator : SY/MD
Sample : N3392-03 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXEX8

Quant Time: Jun 23 02:06:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 23 02:04:00 2022
Response via : Initial Calibration

