

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
 Data File : VU049358.D
 Acq On : 22 Jun 2022 17:43
 Operator : SY/MD
 Sample : N3392-12 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXEY7

Quant Time: Jun 23 02:08:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 23 02:04:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	232158	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	212458	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	91481	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	46674	32.635	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	65.260%	
7) Chloroethane-d5	1.903	69	36955	35.191	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	70.380%	
11) 1,1-Dichloroethene-d2	2.562	63	68308	25.311	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	50.620%#	
21) 2-Butanone-d5	4.623	46	180482	115.099	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	115.100%	
24) Chloroform-d	5.063	84	133742	43.836	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.680%	
26) 1,2-Dichloroethane-d4	5.700	65	91584	46.333	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	92.660%	
32) Benzene-d6	5.726	84	234840	44.576	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	89.160%	
36) 1,2-Dichloropropane-d6	6.690	67	92364	52.152	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	104.300%	
41) Toluene-d8	7.896	98	197519	42.512	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	85.020%	
43) trans-1,3-Dichloroprop...	8.179	79	38933	46.727	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	93.460%	
47) 2-Hexanone-d5	8.632	63	113830	122.804	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	122.800%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	147379	49.823	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	99.640%	
66) 1,2-Dichlorobenzene-d4	12.192	152	76637	48.147	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	96.300%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
Data File : VU049358.D
Acq On : 22 Jun 2022 17:43
Operator : SY/MD
Sample : N3392-12 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXEY7

Quant Time: Jun 23 02:08:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 23 02:04:00 2022
Response via : Initial Calibration

