

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
 Data File : VU049344.D
 Acq On : 22 Jun 2022 12:14
 Operator : SY/MD
 Sample : N3393-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXEZ7

Quant Time: Jun 23 02:05:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 23 02:04:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	240229	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	220333	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	99318	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	58592	39.591	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	79.180%	
7) Chloroethane-d5	1.906	69	49581	45.629	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	91.260%	
11) 1,1-Dichloroethene-d2	2.565	63	85818	30.731	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	61.460%	
21) 2-Butanone-d5	4.623	46	187516	115.567	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	115.570%	
24) Chloroform-d	5.063	84	148257	46.960	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	93.920%	
26) 1,2-Dichloroethane-d4	5.700	65	98311	48.066	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.140%	
32) Benzene-d6	5.726	84	272220	49.825	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.640%	
36) 1,2-Dichloropropane-d6	6.690	67	103038	56.100	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	112.200%	
41) Toluene-d8	7.896	98	234788	48.728	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	97.460%	
43) trans-1,3-Dichloroprop...	8.179	79	42667	49.379	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	98.760%	
47) 2-Hexanone-d5	8.632	63	115745	120.407	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	120.410%	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	157955	51.490	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	102.980%	
66) 1,2-Dichlorobenzene-d4	12.192	152	88753	51.359	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	102.720%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
Data File : VU049344.D
Acq On : 22 Jun 2022 12:14
Operator : SY/MD
Sample : N3393-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXEZ7

Quant Time: Jun 23 02:05:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 23 02:04:00 2022
Response via : Initial Calibration

