

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
 Data File : VU049347.D
 Acq On : 22 Jun 2022 13:25
 Operator : SY/MD
 Sample : N3392-01 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXEX6

Quant Time: Jun 23 02:05:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 23 02:04:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	244180	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	223622	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	99178	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	57005	37.896	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.800%
7) Chloroethane-d5	1.906	69	48456	43.872	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.740%
11) 1,1-Dichloroethene-d2	2.565	63	85011	29.949	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.900%#
21) 2-Butanone-d5	4.623	46	184588	111.922	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.920%
24) Chloroform-d	5.063	84	145992	45.495	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.980%
26) 1,2-Dichloroethane-d4	5.700	65	98598	47.426	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.860%
32) Benzene-d6	5.726	84	268664	48.451	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.900%
36) 1,2-Dichloropropane-d6	6.690	67	101936	54.683	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	109.360%
41) Toluene-d8	7.896	98	234186	47.888	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.780%
43) trans-1,3-Dichloroprop...	8.179	79	43452	49.547	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.100%
47) 2-Hexanone-d5	8.632	63	116186	119.088	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	119.090%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	156637	50.310	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.620%
66) 1,2-Dichlorobenzene-d4	12.192	152	87897	50.936	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
Data File : VU049347.D
Acq On : 22 Jun 2022 13:25
Operator : SY/MD
Sample : N3392-01 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXEX6

Quant Time: Jun 23 02:05:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 23 02:04:00 2022
Response via : Initial Calibration

