

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
 Data File : VU049361.D
 Acq On : 22 Jun 2022 18:54
 Operator : SY/MD
 Sample : N3392-16 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXEZ1

Quant Time: Jun 23 02:08:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 23 02:04:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	225313	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	212345	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	95536	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	41555	29.938	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	59.880%#	
7) Chloroethane-d5	1.896	69	37533	36.828	ug/L	-0.01
Spiked Amount	50.000	Range 70 - 130	Recovery	=	73.660%	
11) 1,1-Dichloroethene-d2	2.562	63	60045	22.925	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	45.860%#	
21) 2-Butanone-d5	4.617	46	166816	109.616	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	109.620%	
24) Chloroform-d	5.060	84	120775	40.788	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	81.580%	
26) 1,2-Dichloroethane-d4	5.700	65	81623	42.548	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	85.100%	
32) Benzene-d6	5.726	84	209885	39.861	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	79.720%	
36) 1,2-Dichloropropane-d6	6.687	67	84941	47.986	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	95.980%	
41) Toluene-d8	7.896	98	177422	38.207	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	76.420%#	
43) trans-1,3-Dichloroprop...	8.179	79	34491	41.418	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	82.840%	
47) 2-Hexanone-d5	8.632	63	104667	112.979	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	112.980%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	145317	49.152	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	98.300%	
66) 1,2-Dichlorobenzene-d4	12.192	152	75614	45.488	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	90.980%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
Data File : VU049361.D
Acq On : 22 Jun 2022 18:54
Operator : SY/MD
Sample : N3392-16 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXEZ1

Quant Time: Jun 23 02:08:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 23 02:04:00 2022
Response via : Initial Calibration

