

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU062222\  
 Data File : VU049363.D  
 Acq On : 22 Jun 2022 19:41  
 Operator : SY/MD  
 Sample : N3392-18 10X  
 Misc : 5.0mL/MSVOA\_U/WATER  
 ALS Vial : 26 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 EXEZ3

Quant Time: Jun 23 02:09:05 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMULM061022WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Thu Jun 23 02:04:00 2022  
 Response via : Initial Calibration

| Compound                      | R.T.    | QIon           | Response | Conc    | Units    | Dev(Min) |
|-------------------------------|---------|----------------|----------|---------|----------|----------|
| -----                         |         |                |          |         |          |          |
| Internal Standards            |         |                |          |         |          |          |
| 1) 1,4-Difluorobenzene        | 6.247   | 114            | 230093   | 50.000  | ug/L     | 0.00     |
| 28) Chlorobenzene-d5          | 9.417   | 117            | 214886   | 50.000  | ug/L     | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.812  | 152            | 95657    | 50.000  | ug/L     | 0.00     |
| System Monitoring Compounds   |         |                |          |         |          |          |
| 4) Vinyl Chloride-d3          | 1.597   | 65             | 45694    | 32.236  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 60 - 135 | Recovery | =       | 64.480%  |          |
| 7) Chloroethane-d5            | 1.906   | 69             | 40640    | 39.048  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 70 - 130 | Recovery | =       | 78.100%  |          |
| 11) 1,1-Dichloroethene-d2     | 2.565   | 63             | 64694    | 24.187  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 60 - 125 | Recovery | =       | 48.380%# |          |
| 21) 2-Butanone-d5             | 4.620   | 46             | 165612   | 106.564 | ug/L     | 0.00     |
| Spiked Amount                 | 100.000 | Range 40 - 130 | Recovery | =       | 106.560% |          |
| 24) Chloroform-d              | 5.060   | 84             | 121180   | 40.075  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 70 - 125 | Recovery | =       | 80.140%  |          |
| 26) 1,2-Dichloroethane-d4     | 5.700   | 65             | 83389    | 42.566  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 70 - 125 | Recovery | =       | 85.140%  |          |
| 32) Benzene-d6                | 5.726   | 84             | 216498   | 40.630  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 70 - 125 | Recovery | =       | 81.260%  |          |
| 36) 1,2-Dichloropropane-d6    | 6.690   | 67             | 84707    | 47.288  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 70 - 120 | Recovery | =       | 94.580%  |          |
| 41) Toluene-d8                | 7.896   | 98             | 180625   | 38.437  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 80 - 120 | Recovery | =       | 76.880%# |          |
| 43) trans-1,3-Dichloroprop... | 8.179   | 79             | 34525    | 40.969  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 60 - 125 | Recovery | =       | 81.940%  |          |
| 47) 2-Hexanone-d5             | 8.629   | 63             | 101257   | 108.006 | ug/L     | 0.00     |
| Spiked Amount                 | 100.000 | Range 45 - 130 | Recovery | =       | 108.010% |          |
| 56) 1,1,2,2-Tetrachloroeth... | 10.755  | 84             | 140091   | 46.824  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 65 - 120 | Recovery | =       | 93.640%  |          |
| 66) 1,2-Dichlorobenzene-d4    | 12.192  | 152            | 75415    | 45.311  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 80 - 120 | Recovery | =       | 90.620%  |          |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU062222\  
Data File : VU049363.D  
Acq On : 22 Jun 2022 19:41  
Operator : SY/MD  
Sample : N3392-18 10X  
Misc : 5.0mL/MSVOA\_U/WATER  
ALS Vial : 26 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
EXEZ3

Quant Time: Jun 23 02:09:05 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMULM061022WMA.M  
Quant Title : VOC Analysis  
QLast Update : Thu Jun 23 02:04:00 2022  
Response via : Initial Calibration

