

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
 Data File : VU049370.D
 Acq On : 23 Jun 2022 11:38
 Operator : SY/MD
 Sample : N3392-21 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXEZ6

Quant Time: Jun 24 01:11:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 24 01:09:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	201349	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	180056	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	76236	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	40400	32.570	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	65.140%	
7) Chloroethane-d5	1.906	69	37240	40.889	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	81.780%	
11) 1,1-Dichloroethene-d2	2.565	63	55594	23.752	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	47.500%#	
21) 2-Butanone-d5	4.623	46	174897	128.604	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	128.600%	
24) Chloroform-d	5.063	84	115608	43.690	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.380%	
26) 1,2-Dichloroethane-d4	5.703	65	81300	47.424	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.840%	
32) Benzene-d6	5.726	84	195239	43.728	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.460%	
36) 1,2-Dichloropropane-d6	6.690	67	80916	53.910	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	107.820%	
41) Toluene-d8	7.896	98	155648	39.529	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	79.060%#	
43) trans-1,3-Dichloroprop...	8.179	79	30912	43.777	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	87.560%	
47) 2-Hexanone-d5	8.632	63	104386	132.882	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	132.880%#	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	135726	54.141	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	108.280%	
66) 1,2-Dichlorobenzene-d4	12.192	152	63539	47.901	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	95.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
Data File : VU049370.D
Acq On : 23 Jun 2022 11:38
Operator : SY/MD
Sample : N3392-21 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXEZ6

Quant Time: Jun 24 01:11:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 24 01:09:54 2022
Response via : Initial Calibration

