

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
 Data File : VU049371.D
 Acq On : 23 Jun 2022 12:01
 Operator : SY/MD
 Sample : PB145709TB 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VLEB709

Quant Time: Jun 24 01:11:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 24 01:09:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	232770	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	216479	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	96491	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	71895	50.137	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 100.280%	
7) Chloroethane-d5	1.903	69	56573	53.732	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 107.460%	
11) 1,1-Dichloroethene-d2	2.565	63	96596	35.699	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 71.400%	
21) 2-Butanone-d5	4.617	46	175160	111.411	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 111.410%	
24) Chloroform-d	5.060	84	152255	49.772	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 99.540%	
26) 1,2-Dichloroethane-d4	5.700	65	101025	50.975	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 101.960%	
32) Benzene-d6	5.726	84	282657	52.656	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 105.320%	
36) 1,2-Dichloropropane-d6	6.687	67	104299	57.797	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 115.600%	
41) Toluene-d8	7.896	98	246921	52.158	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 104.320%	
43) trans-1,3-Dichloroprop...	8.179	79	41514	48.900	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 97.800%	
47) 2-Hexanone-d5	8.629	63	108182	114.543	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 114.540%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	156343	51.872	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 103.740%	
66) 1,2-Dichlorobenzene-d4	12.192	152	92810	55.281	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 110.560%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
Data File : VU049371.D
Acq On : 23 Jun 2022 12:01
Operator : SY/MD
Sample : PB145709TB 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VLEB709

Quant Time: Jun 24 01:11:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 24 01:09:54 2022
Response via : Initial Calibration

