

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
 Data File : VU049372.D
 Acq On : 23 Jun 2022 12:25
 Operator : SY/MD
 Sample : N3392-20RE 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXEZ5RE

Quant Time: Jun 24 01:11:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 24 01:09:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.246	114	184052	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	171711	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	74261	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	36708	32.375	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	64.740%	
7) Chloroethane-d5	1.899	69	33276	39.970	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	79.940%	
11) 1,1-Dichloroethene-d2	2.562	63	49620	23.192	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	46.380%#	
21) 2-Butanone-d5	4.616	46	166499	133.934	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	133.930%#	
24) Chloroform-d	5.060	84	108898	45.022	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	90.040%	
26) 1,2-Dichloroethane-d4	5.700	65	79007	50.418	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.840%	
32) Benzene-d6	5.726	84	183876	43.185	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	86.360%	
36) 1,2-Dichloropropane-d6	6.690	67	76533	53.468	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	106.940%	
41) Toluene-d8	7.896	98	146109	38.910	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	77.820%#	
43) trans-1,3-Dichloroprop...	8.179	79	30281	44.967	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	89.940%	
47) 2-Hexanone-d5	8.632	63	102440	136.742	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	136.740%#	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	138667	58.002	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	116.000%	
66) 1,2-Dichlorobenzene-d4	12.192	152	62923	48.698	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	97.400%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062222\
Data File : VU049372.D
Acq On : 23 Jun 2022 12:25
Operator : SY/MD
Sample : N3392-20RE 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXEZ5RE

Quant Time: Jun 24 01:11:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 24 01:09:54 2022
Response via : Initial Calibration

