

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062223\
 Data File : VU054659.D
 Acq On : 22 Jun 2023 16:26
 Operator : MD/SY
 Sample : 03384-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YBW91

Quant Time: Jun 23 04:35:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 23 04:31:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	295062	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	285113	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	127080	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	100337	5.343	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.800%
7) Chloroethane-d5	1.914	69	83577	4.754	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.000%
11) 1,1-Dichloroethene-d2	2.563	65	40327	4.808	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.200%
20) 2-Butanone-d5	4.627	46	244577	46.325	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.640%
24) Chloroform-d	5.058	84	166004	4.229	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.600%
26) 1,2-Dichloroethane-d4	5.698	65	84971	4.441	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.800%
32) Benzene-d6	5.724	84	330301	4.167	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.400%
36) 1,2-Dichloropropane-d6	6.688	67	111724	4.221	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.400%
41) Toluene-d8	7.894	98	275033	3.910	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.200%
43) trans-1,3-Dichloroprop...	8.177	79	38845	4.220	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.400%
46) 2-Hexanone-d5	8.630	63	161495	41.821	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.640%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	81420	4.125	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	94348	4.402	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062223\
Data File : VU054659.D
Acq On : 22 Jun 2023 16:26
Operator : MD/SY
Sample : 03384-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YBW91

Quant Time: Jun 23 04:35:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 23 04:31:37 2023
Response via : Initial Calibration

