

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062318\
 Data File : VU024844.D
 Acq On : 22 Jun 2018 20:14
 Operator : MD/SY
 Sample : J3561-02DL 20X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S32-0624DL

Quant Time: Jun 23 04:03:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	202868	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	308966	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	288673	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	155590	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	145238	43.85	ug/l	0.00
Spiked Amount			Recovery	=	87.70%	
35) Dibromofluoromethane	4.89	113	119857	46.74	ug/l	0.00
Spiked Amount			Recovery	=	93.48%	
50) Toluene-d8	7.57	98	454916	48.64	ug/l	0.00
Spiked Amount			Recovery	=	97.28%	
62) 4-Bromofluorobenzene	10.31	95	167668	45.15	ug/l	0.00
Spiked Amount			Recovery	=	90.30%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	135487	50.64	ug/l	98
21) trans-1,2-Dichloroethene	2.99	96	3728	1.47	ug/l #	82
27) cis-1,2-Dichloroethene	4.23	96	64568	22.07	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062318\
Data File : VU024844.D
Acq On : 22 Jun 2018 20:14
Operator : MD/SY
Sample : J3561-02DL 20X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S32-0624DL

Quant Time: Jun 23 04:03:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

