

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062318\
 Data File : VU024846.D
 Acq On : 22 Jun 2018 21:03
 Operator : MD/SY
 Sample : J3517-05
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S39-ER-2018-2

Quant Time: Jun 23 04:07:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	199787	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	306709	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	283614	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	153657	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	146772	45.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.00%	
35) Dibromofluoromethane	4.89	113	118652	46.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.22%	
50) Toluene-d8	7.57	98	439743	47.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.72%	
62) 4-Bromofluorobenzene	10.31	95	162896	44.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.38%	
Target Compounds						
16) Acetone	2.32	43	6734	3.84	ug/l	98
20) Methylene Chloride	2.71	84	101926	37.07	ug/l	96
25) 2-Butanone	4.27	43	27241	11.57	ug/l	98
29) Tetrahydrofuran	4.63	42	905438	621.14	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062318\
Data File : VU024846.D
Acq On : 22 Jun 2018 21:03
Operator : MD/SY
Sample : J3517-05
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S39-ER-2018-2

Quant Time: Jun 23 04:07:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

