

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062318\
 Data File : VU024855.D
 Acq On : 23 Jun 2018 00:41
 Operator : MD/SY
 Sample : J3662-07
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021MW038-180619

Quant Time: Jun 23 04:39:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	197633	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	303100	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	278742	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	151784	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	146815	45.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.00%	
35) Dibromofluoromethane	4.89	113	117136	46.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.14%	
50) Toluene-d8	7.57	98	441108	48.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.14%	
62) 4-Bromofluorobenzene	10.31	95	161279	44.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.54%	
Target Compounds						
4) Vinyl Chloride	1.40	62	4277	1.64	ug/l	95
27) cis-1,2-Dichloroethene	4.23	96	29934	10.51	ug/l	86
64) Tetrachloroethene	8.23	164	10907	4.12	ug/l	95
65) Chlorobenzene	9.12	112	16366	2.30	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062318\
Data File : VU024855.D
Acq On : 23 Jun 2018 00:41
Operator : MD/SY
Sample : J3662-07
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021MW038-180619

Quant Time: Jun 23 04:39:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

