

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062318\
 Data File : VU024856.D
 Acq On : 23 Jun 2018 01:05
 Operator : MD/SY
 Sample : J3662-08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021MW348-180619

Quant Time: Jun 23 04:41:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	206561	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	312964	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	291125	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	154367	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	151583	44.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.90%	
35) Dibromofluoromethane	4.89	113	122028	46.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.96%	
50) Toluene-d8	7.57	98	453034	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
62) 4-Bromofluorobenzene	10.31	95	165770	44.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.14%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	1764	0.65	ug/l	# 85
27) cis-1,2-Dichloroethene	4.23	96	25575	8.59	ug/l	89
44) Trichloroethene	6.19	130	2415	0.81	ug/l	93
64) Tetrachloroethene	8.23	164	3774	1.36	ug/l	94
65) Chlorobenzene	9.12	112	2292	0.31	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062318\
Data File : VU024856.D
Acq On : 23 Jun 2018 01:05
Operator : MD/SY
Sample : J3662-08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021MW348-180619

Quant Time: Jun 23 04:41:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

