

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062419\
 Data File : VU033031.D
 Acq On : 25 Jun 2019 01:12
 Operator : JC/SP
 Sample : K3463-11DL 10X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETF14DL

Quant Time: Jun 25 07:59:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 25 06:54:49 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	197348	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	194338	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	92118	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	62906	46.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.28%
7) Chloroethane-d5	1.67	69	52792	49.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.16%
11) 1,1-Dichloroethene-d2	2.26	63	99037	36.45	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.90%
21) 2-Butanone-d5	4.17	46	125227	111.55	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.55%
24) Chloroform-d	4.64	84	130770	49.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.76%
26) 1,2-Dichloroethane-d4	5.30	65	94462	54.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.18%
32) Benzene-d6	5.33	84	256646	50.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.32%
36) 1,2-Dichloropropane-d6	6.32	67	83737	51.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.78%
41) Toluene-d8	7.56	98	231863	48.87	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.74%
43) trans-1,3-Dichloropropene-	7.85	79	39035	46.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.00%
47) 2-Hexanone-d5	8.31	63	72867	96.70	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.70%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	119021	47.71	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.42%
64) 1,2-Dichlorobenzene-d4	11.85	152	90092	50.54	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.08%

Target Compounds

					Ovalue
5) Vinyl chloride	1.40	62	78287	43.685	ug/L 100
17) trans-1,2-Dichloroethene	2.97	96	520	0.386	ug/L 95
20) cis-1,2-Dichloroethene	4.22	96	119023	81.070	ug/L 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062419\
Data File : VU033031.D
Acq On : 25 Jun 2019 01:12
Operator : JC/SP
Sample : K3463-11DL 10X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETF14DL

Quant Time: Jun 25 07:59:10 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 25 06:54:49 2019
Response via : Initial Calibration

