

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062419\
 Data File : VU033011.D
 Acq On : 24 Jun 2019 17:33
 Operator : JC/SP
 Sample : K3496-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETF64

Quant Time: Jun 25 07:05:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 25 06:54:49 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	194017	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	188055	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	90449	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	63438	47.84	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.68%
7) Chloroethane-d5	1.68	69	53718	50.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.60%
11) 1,1-Dichloroethene-d2	2.27	63	99862	37.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.78%
21) 2-Butanone-d5	4.17	46	122151	110.68	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.68%
24) Chloroform-d	4.64	84	131848	51.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.30%
26) 1,2-Dichloroethane-d4	5.30	65	91343	53.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.40%
32) Benzene-d6	5.33	84	256603	52.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.68%
36) 1,2-Dichloropropane-d6	6.32	67	82179	52.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.24%
41) Toluene-d8	7.56	98	233316	50.82	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.64%
43) trans-1,3-Dichloropropene-	7.85	79	40954	50.41	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.82%
47) 2-Hexanone-d5	8.31	63	73004	100.12	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.12%
57) 1,1,2,2-Tetrachloroethane-	10.42	84	120200	49.79	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.58%
64) 1,2-Dichlorobenzene-d4	11.85	152	90458	51.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.36%


Target Compounds					Ovalue
13) Acetone	2.32	43	9223	6.051 ug/L	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062419\
Data File : VU033011.D
Acq On : 24 Jun 2019 17:33
Operator : JC/SP
Sample : K3496-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETF64

Quant Time: Jun 25 07:05:52 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 25 06:54:49 2019
Response via : Initial Calibration





#13

Acetone

Concen: 6.051 ug/L

RT: 2.32 min Scan# 382

Delta R.T. 0.00 min

Lab File: VU033011.D

Acq: 24 Jun 2019 17:33

Instrument :
MSVOA_U
ClientSampleId :
ETF64

Tot Ion: 43 Resp: 9223

Ion Ratio Lower Upper

43 100

58 26.3 0.0 60.6

