

Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062420\
 Data File : VU039077.D
 Acq On : 24 Jun 2020 22:00
 Operator : SY/MD
 Sample : L3110-14
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
ClientSampleId :
 F9F41

Manual Integrations
APPROVED

MMDadoda
 6/25/2020 6:08:46 PM

Quant Time: Jun 25 02:36:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 25 02:33:05 2020
 Response via : Initial Calibration



TIC: VU039077.D

(13) Acetone (T)

2.684min (+0.013) 0.87ug/L

response 3900

Ion	Exp%	Act%
43.05	100	100
58.05	32.20	19.77
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062420\
 Data File : VU039077.D
 Acq On : 24 Jun 2020 22:00
 Operator : SY/MD
 Sample : L3110-14
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9F41

Manual Integrations
 APPROVED

MMDadoda
 6/25/2020 6:08:46 PM

Quant Time: Jun 25 02:36:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 25 02:33:05 2020
 Response via : Initial Calibration



TIC: VU039077.D

(13) Acetone (T)

2.684min (+0.013) 1.63ug/L m

response 7312

Ion	Exp%	Act%
43.05	100	100
58.05	32.20	10.54
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062420\
 Data File : VU039077.D
 Acq On : 24 Jun 2020 22:00
 Operator : SY/MD
 Sample : L3110-14
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9F41

Manual Integrations
 APPROVED

MMDadoda
 6/25/2020 6:08:46 PM

Quant Time: Jun 25 03:08:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 25 02:33:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	249321	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.43	117	255653	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	108718	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.61	65	91706	5.02	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery = 100.40%			
7) Chloroethane-d5	1.93	69	84563	5.60	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery = 112.00%			
11) 1,1-Dichloroethene-d2	2.59	63	138044	3.93	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery = 78.60%			
20) 2-Butanone-d5	4.68	46	423913	64.34	ug/L	0.00
Spiked Amount 50.000	Range 40 - 130		Recovery = 128.68%			
24) Chloroform-d	5.10	84	198900	5.54	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery = 110.80%			
26) 1,2-Dichloroethane-d4	5.74	65	123785	5.60	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery = 112.00%			
32) Benzene-d6	5.76	84	376742	5.31	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery = 106.20%			
36) 1,2-Dichloropropane-d6	6.72	67	126824	5.32	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery = 106.40%			
41) Toluene-d8	7.92	98	294542	4.97	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery = 99.40%			
43) trans-1,3-Dichloropropene-	8.20	79	51788	5.00	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery = 100.00%			
46) 2-Hexanone-d5	8.66	63	318712	59.64	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery = 119.28%			
57) 1,1,2,2-Tetrachloroethane-	10.77	84	116990	5.28	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery = 105.60%			
64) 1,2-Dichlorobenzene-d4	12.20	152	116097	5.73	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery = 114.60%			
Target Compounds						
13) Acetone	2.68	43	7173m	1.603	ug/L	Ovalue
25) Chloroform	5.12	83	88093	2.387	ug/L	98
38) Bromodichloromethane	7.13	83	21686	0.799	ug/L	96
49) Dibromochloromethane	8.83	129	5539	0.315	ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062420\
Data File : VU039077.D
Acq On : 24 Jun 2020 22:00
Operator : SY/MD
Sample : L3110-14
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9F41

Manual Integrations
APPROVED

MMDadoda
6/25/2020 6:08:46 PM

Quant Time: Jun 25 03:08:57 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\SOMUTR062320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 25 02:33:05 2020
Response via : Initial Calibration

