

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062420\
 Data File : VU039092.D
 Acq On : 25 Jun 2020 03:45
 Operator : SY/MD
 Sample : L3110-08
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9F23

Quant Time: Jun 25 06:56:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 25 06:40:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	252508	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	254098	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	110278	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	95286	5.15	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.00%
7) Chloroethane-d5	1.93	69	88171	5.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.40%
11) 1,1-Dichloroethene-d2	2.59	63	144651	4.07	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.40%
20) 2-Butanone-d5	4.67	46	412796	61.86	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.72%
24) Chloroform-d	5.10	84	201027	5.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.60%
26) 1,2-Dichloroethane-d4	5.74	65	125844	5.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.60%
32) Benzene-d6	5.76	84	392543	5.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.40%
36) 1,2-Dichloropropane-d6	6.72	67	129957	5.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.80%
41) Toluene-d8	7.92	98	302504	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.60%
43) trans-1,3-Dichloropropene-	8.20	79	51005	4.96	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.20%
46) 2-Hexanone-d5	8.65	63	323556	60.92	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	121.84%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	117573	5.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.80%
64) 1,2-Dichlorobenzene-d4	12.20	152	120692	5.87	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.40%


Target Compounds					Ovalue
47) Tetrachloroethene	8.57	164	252906	18.171 ug/L	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062420\
Data File : VU039092.D
Acq On : 25 Jun 2020 03:45
Operator : SY/MD
Sample : L3110-08
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9F23

Quant Time: Jun 25 06:56:53 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 25 06:40:48 2020
Response via : Initial Calibration





#47

Tetrachloroethene

Concen: 18.171 ug/L

RT: 8.57 min Scan# 2249

Delta R.T. -0.00 min

Lab File: VU039092.D

Acq: 25 Jun 2020 03:45

Instrument :

MSVOA_U

ClientSampleId :

F9F23

Tot Ion:164 Resp: 252906


Ion Ratio Lower Upper

164 100

129 97.0 71.3 132.5

131 92.3 68.3 126.8

166 121.5 90.8 168.6



Abundance Ion 163.90 (163.60 to 164.60): V

Ion 128.95 (128.65 to 129.65): V

Ion 130.95 (130.65 to 131.65): V

Ion 165.90 (165.60 to 166.60): V

