

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062422\
 Data File : VU049393.D
 Acq On : 24 Jun 2022 17:30
 Operator : SY/MD
 Sample : N3400-22
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Jun 25 00:18:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 25 00:13:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	136275	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	126961	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	60170	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	39816	3.344	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.800%
7) Chloroethane-d5	1.912	69	31392	3.894	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.800%
11) 1,1-Dichloroethene-d2	2.565	65	17257	3.603	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.000%
20) 2-Butanone-d5	4.633	46	165300	51.991	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.980%
24) Chloroform-d	5.063	84	82656	4.424	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
26) 1,2-Dichloroethane-d4	5.703	65	47090	4.595	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.800%
32) Benzene-d6	5.726	84	157014	4.105	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.200%
36) 1,2-Dichloropropane-d6	6.690	67	55434	4.270	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.400%
41) Toluene-d8	7.896	98	137292	3.986	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.800%
43) trans-1,3-Dichloroprop...	8.179	79	20633	3.578	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	71.600%
46) 2-Hexanone-d5	8.632	63	116229	49.961	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.920%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	49773	4.789	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	12.192	152	52940	4.585	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062422\
Data File : VU049393.D
Acq On : 24 Jun 2022 17:30
Operator : SY/MD
Sample : N3400-22
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Jun 25 00:18:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 25 00:13:49 2022
Response via : Initial Calibration

