




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062422\
Data File : VU049379.D
Acq On : 24 Jun 2022 11:24
Operator : SY/MD
Sample : N3400-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AD3

Manual IntegrationsAPPROVED

Quant Time: Jun 25 00:15:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 25 00:13:49 2022
Response via : Initial Calibration

Reviewed By :John Carlone 06/28/2022
Supervised By :Mahesh Dadoda 06/28/2022



TIC: VU049379.D\data.ms

(13) Acetone (T)

2.645min (-0.010) 5.61 ug/L

response 7889

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	32.90	12.84
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062422\
 Data File : VU049379.D
 Acq On : 24 Jun 2022 11:24
 Operator : SY/MD
 Sample : N3400-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AD3

Manual IntegrationsAPPROVED

Quant Time: Jun 25 00:15:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 25 00:13:49 2022
 Response via : Initial Calibration

Reviewed By :John Carlone 06/28/2022
 Supervised By :Mahesh Dadoda 06/28/2022



TIC: VU049379.D\data.ms

(13) Acetone (T)

2.645min (-0.010) 6.70 ug/L m

response 9425

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	32.90	10.75
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062422\
 Data File : VU049379.D
 Acq On : 24 Jun 2022 11:24
 Operator : SY/MD
 Sample : N3400-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AD3

Manual IntegrationsAPPROVED

Quant Time: Jun 25 00:15:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 25 00:13:49 2022
 Response via : Initial Calibration

Reviewed By : John Carlone 06/28/2022
 Supervised By : Mahesh Dadoda 06/28/2022

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	6.250	114		121877	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117		109808	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152		50504	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	1.597	65		32348	3.038	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	60.800%	
7) Chloroethane-d5	1.912	69		25600	3.551	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	71.000%	
11) 1,1-Dichloroethene-d2	2.568	65		14810	3.458	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.200%	
20) 2-Butanone-d5	4.639	46		93018	32.713	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	65.420%	
24) Chloroform-d	5.063	84		73430	4.395	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.800%	
26) 1,2-Dichloroethane-d4	5.703	65		41287	4.505	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.000%	
32) Benzene-d6	5.729	84		140262	4.240	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.800%	
36) 1,2-Dichloropropane-d6	6.690	67		49597	4.417	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.400%	
41) Toluene-d8	7.899	98		124186	4.169	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.400%	
43) trans-1,3-Dichloroprop...	8.182	79		14052	2.817	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	56.400%	
46) 2-Hexanone-d5	8.636	63		69246	34.415	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	68.820%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84		40512	4.507	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.200%	
66) 1,2-Dichlorobenzene-d4	12.195	152		44979	4.641	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.800%	
Target Compounds							
13) Acetone	2.645	43		9425m	6.697	ug/L	
42) Toluene	7.970	91		20092	0.580	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062422\
Data File : VU049379.D
Acq On : 24 Jun 2022 11:24
Operator : SY/MD
Sample : N3400-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AD3

Manual IntegrationsAPPROVED

Quant Time: Jun 25 00:15:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 25 00:13:49 2022
Response via : Initial Calibration

Reviewed By :John Carlone 06/28/2022
Supervised By :Mahesh Dadoda 06/28/2022

