

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062518\
 Data File : VU024877.D
 Acq On : 25 Jun 2018 16:08
 Operator : MD/SY
 Sample : J3662-09DL 100X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021MW346-180619DL

Quant Time: Jun 26 07:28:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	201870	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	309636	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	287432	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	155334	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	150471	45.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.32%	
35) Dibromofluoromethane	4.89	113	120105	46.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.48%	
50) Toluene-d8	7.57	98	451798	48.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.40%	
62) 4-Bromofluorobenzene	10.31	95	166461	44.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.46%	
Target Compounds						
4) Vinyl Chloride	1.40	62	39400	14.80	ug/l	Qvalue 100
27) cis-1,2-Dichloroethene	4.23	96	129744	44.58	ug/l	89
44) Trichloroethene	6.19	130	2330	0.79	ug/l	79
65) Chlorobenzene	9.12	112	3728	0.51	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062518\
Data File : VU024877.D
Acq On : 25 Jun 2018 16:08
Operator : MD/SY
Sample : J3662-09DL 100X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021MW346-180619DL

Quant Time: Jun 26 07:28:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

