

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062518\
 Data File : VU024878.D
 Acq On : 25 Jun 2018 16:32
 Operator : MD/SY
 Sample : J3662-10
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021MW173-180619

Quant Time: Jun 26 12:10:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	198150	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	303543	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	280211	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	148652	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	148086	45.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.56%	
35) Dibromofluoromethane	4.89	113	119108	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
50) Toluene-d8	7.57	98	440157	47.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.80%	
62) 4-Bromofluorobenzene	10.31	95	162316	44.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.98%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	5194	1.99	ug/l	97
27) cis-1,2-Dichloroethene	4.23	96	11867	4.15	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062518\
Data File : VU024878.D
Acq On : 25 Jun 2018 16:32
Operator : MD/SY
Sample : J3662-10
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021MW173-180619

Quant Time: Jun 26 12:10:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

