

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062518\
 Data File : VU024880.D
 Acq On : 25 Jun 2018 17:20
 Operator : MD/SY
 Sample : J3605-01ME
 Misc : 6.02g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S-1(5-5.5)ME

Quant Time: Jun 26 12:13:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	210948	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	320239	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	310784	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	182349	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	156609	45.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.96%	
35) Dibromofluoromethane	4.89	113	123764	46.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.14%	
50) Toluene-d8	7.57	98	469054	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
62) 4-Bromofluorobenzene	10.32	95	196251	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
Target Compounds						
64) Tetrachloroethene	8.23	164	53952	18.26	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062518\
Data File : VU024880.D
Acq On : 25 Jun 2018 17:20
Operator : MD/SY
Sample : J3605-01ME
Misc : 6.02g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S-1(5-5.5)ME

Quant Time: Jun 26 12:13:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

