

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062518\
 Data File : VU024883.D
 Acq On : 25 Jun 2018 18:33
 Operator : MD/SY
 Sample : J3605-04ME
 Misc : 6.01g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S-3(8.5-9)ME

Quant Time: Jun 26 12:16:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	208760	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	322380	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.10	117	317625	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	189235	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	156574	45.94	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.88%	
35) Dibromofluoromethane	4.89	113	124493	46.53	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	93.06%	
50) Toluene-d8	7.57	98	473228	48.49	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.98%	
62) 4-Bromofluorobenzene	10.31	95	200772	51.82	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	103.64%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
64) Tetrachloroethene	8.23	164	1263486	418.49	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062518\
Data File : VU024883.D
Acq On : 25 Jun 2018 18:33
Operator : MD/SY
Sample : J3605-04ME
Misc : 6.01µ/5.0mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S-3(8.5-9)ME

Quant Time: Jun 26 12:16:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

