

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062518\
 Data File : VU024899.D
 Acq On : 26 Jun 2018 01:25
 Operator : MD/SY
 Sample : J3669-22ME
 Misc : 6.75u/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB479-25-27-180620ME

Manual Integrations
 APPROVED

apatel
 6/27/2018 11:21:26 AM

Quant Time: Jun 26 09:45:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	230952	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	355100	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.10	117	342192	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	224063	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	175381	46.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.04%	
35) Dibromofluoromethane	4.89	113	138572	47.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.04%	
50) Toluene-d8	7.57	98	526377	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
62) 4-Bromofluorobenzene	10.32	95	248161	58.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.28%	
Target Compounds						
39) Methylcyclohexane	6.41	83	46891	9.01	ug/l	96
65) Chlorobenzene	9.13	112	78728	9.01	ug/l	98
83) tert-Butylbenzene	11.11	119	141844	9.89	ug/l	93
85) sec-Butylbenzene	11.33	105	522095	29.08	ug/l	98
89) n-Butylbenzene	11.90	91	457441m	32.04	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062518\
Data File : VU024899.D
Acq On : 26 Jun 2018 01:25
Operator : MD/SY
Sample : J3669-22ME
Misc : 6.75µ/5.0mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB479-25-27-180620ME

Manual Integrations
APPROVED

apatel
6/27/2018 11:21:26 AM

Quant Time: Jun 26 09:45:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

