

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062518\
 Data File : VU024900.D
 Acq On : 26 Jun 2018 01:49
 Operator : MD/SY
 Sample : J3669-27ME
 Misc : 6.20g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB474-25-27-180620ME

Quant Time: Jun 26 08:18:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	252482	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	382807	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.10	117	362607	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	217678	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	183223	44.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.90%	
35) Dibromofluoromethane	4.89	113	145760	45.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.76%	
50) Toluene-d8	7.57	98	545260	47.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.10%	
62) 4-Bromofluorobenzene	10.32	95	227487	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
Target Compounds						
65) Chlorobenzene	9.13	112	45254	4.89	ug/l	97
83) tert-Butylbenzene	11.10	119	15580	1.12	ug/l	92
85) sec-Butylbenzene	11.33	105	38954	2.23	ug/l	95
91) 1,2-Dichlorobenzene	11.89	146	9600	1.23	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062518\
Data File : VU024900.D
Acq On : 26 Jun 2018 01:49
Operator : MD/SY
Sample : J3669-27ME
Misc : 6.20µ/5.0mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB474-25-27-180620ME

Quant Time: Jun 26 08:18:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

