

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062518\
 Data File : VU024902.D
 Acq On : 26 Jun 2018 02:38
 Operator : MD/SY
 Sample : J3669-07ME 20X
 Misc : 5.94g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB484-35-37-180620ME

Quant Time: Jun 26 08:25:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	228132	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	348867	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	317434	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	173188	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	170510	45.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.56%	
35) Dibromofluoromethane	4.89	113	133061	45.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.92%	
50) Toluene-d8	7.57	98	499276	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
62) 4-Bromofluorobenzene	10.31	95	186697	44.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.04%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.23	96	5688	1.73	ug/l	84
44) Trichloroethene	6.19	130	14661	4.41	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062518\
Data File : VU024902.D
Acq On : 26 Jun 2018 02:38
Operator : MD/SY
Sample : J3669-07ME 20X
Misc : 5.94g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB484-35-37-180620ME

Quant Time: Jun 26 08:25:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

