

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062518\
 Data File : VU024903.D
 Acq On : 26 Jun 2018 03:02
 Operator : MD/SY
 Sample : J3669-08ME 20X
 Misc : 5.92g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB484-37-39-180620ME

Quant Time: Jun 26 08:28:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	227493	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	347426	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	314528	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	170330	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	168811	45.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.90%	
35) Dibromofluoromethane	4.89	113	135077	46.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.70%	
50) Toluene-d8	7.57	98	496297	47.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.38%	
62) 4-Bromofluorobenzene	10.31	95	184006	44.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.12%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.24	96	2927	0.89	ug/l	91
44) Trichloroethene	6.19	130	10076	3.04	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062518\
Data File : VU024903.D
Acq On : 26 Jun 2018 03:02
Operator : MD/SY
Sample : J3669-08ME 20X
Misc : 5.92g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB484-37-39-180620ME

Quant Time: Jun 26 08:28:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

