

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062518\
 Data File : VU024908.D
 Acq On : 26 Jun 2018 05:03
 Operator : MD/SY
 Sample : J3669-17ME 200X
 Misc : 5.73g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP-0483-180620ME

Quant Time: Jun 26 08:43:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	226908	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	341631	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	317977	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	170276	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	169209	45.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.36%	
35) Dibromofluoromethane	4.89	113	134220	47.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.68%	
50) Toluene-d8	7.57	98	495588	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
62) 4-Bromofluorobenzene	10.31	95	184848	45.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.04%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
44) Trichloroethene	6.19	130	18745	5.76	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062518\
Data File : VU024908.D
Acq On : 26 Jun 2018 05:03
Operator : MD/SY
Sample : J3669-17ME 200X
Misc : 5.73q/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP-0483-180620ME

Quant Time: Jun 26 08:43:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

