

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062518\
 Data File : VU024910.D
 Acq On : 26 Jun 2018 05:51
 Operator : MD/SY
 Sample : J3669-19ME 200X
 Misc : 6.02g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB475-27-29-180620ME

Quant Time: Jun 26 08:46:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	221356	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	336526	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	311686	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	164043	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	165381	45.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.54%	
35) Dibromofluoromethane	4.89	113	131007	46.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.82%	
50) Toluene-d8	7.57	98	478693	46.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.98%	
62) 4-Bromofluorobenzene	10.31	95	178101	44.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.06%	

Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.23	96	2623	0.82	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062518\
Data File : VU024910.D
Acq On : 26 Jun 2018 05:51
Operator : MD/SY
Sample : J3669-19ME 200X
Misc : 6.02g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB475-27-29-180620ME

Quant Time: Jun 26 08:46:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

