

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062518\
 Data File : VU024914.D
 Acq On : 26 Jun 2018 07:27
 Operator : MD/SY
 Sample : J3669-25ME 200X
 Misc : 5.96g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021SB479-31-33-180620ME

Manual Integrations
 APPROVED

apatel
 6/27/2018 11:20:41 AM

Quant Time: Jun 26 08:54:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	220885	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	333541	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	310047	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	167342	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	165475	45.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.78%	
35) Dibromofluoromethane	4.89	113	131634	47.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.10%	
50) Toluene-d8	7.57	98	486295	48.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.32%	
62) 4-Bromofluorobenzene	10.31	95	178449	44.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.02%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.23	96	2438m	0.77	ug/l	Qvalue
44) Trichloroethene	6.19	130	6883	2.17	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062518\
Data File : VU024914.D
Acq On : 26 Jun 2018 07:27
Operator : MD/SY
Sample : J3669-25ME 200X
Misc : 5.96g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021SB479-31-33-180620ME

Manual Integrations
APPROVED

apatel
6/27/2018 11:20:41 AM

Quant Time: Jun 26 08:54:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

