

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062520\
 Data File : VU039108.D
 Acq On : 25 Jun 2020 12:11
 Operator : SY/MD
 Sample : VIBLK
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK

Quant Time: Jun 26 01:44:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 26 00:59:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	285732	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	292090	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	123394	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	71833	3.43	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.60%
7) Chloroethane-d5	1.93	69	68669	3.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.40%
11) 1,1-Dichloroethene-d2	2.59	63	108815	2.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.20%#
20) 2-Butanone-d5	4.67	46	352688	46.71	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.42%
24) Chloroform-d	5.10	84	169377	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.40%
26) 1,2-Dichloroethane-d4	5.74	65	106922	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.40%
32) Benzene-d6	5.76	84	305492	3.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	75.40%
36) 1,2-Dichloropropane-d6	6.72	67	106679	3.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.40%
41) Toluene-d8	7.92	98	228753	3.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	67.60%#
43) trans-1,3-Dichloropropene-	8.20	79	43222	3.65	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.00%
46) 2-Hexanone-d5	8.65	63	265832	43.54	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.08%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	100124	3.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.20%
64) 1,2-Dichlorobenzene-d4	12.20	152	98123	4.27	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.40%

Target Compounds					Ovalue
16) Methylene chloride	3.08	84	3946	0.162 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062520\
Data File : VU039108.D
Acq On : 25 Jun 2020 12:11
Operator : SY/MD
Sample : VIBLK
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK

Quant Time: Jun 26 01:44:59 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 26 00:59:08 2020
Response via : Initial Calibration

