

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062520\
 Data File : VU039124.D
 Acq On : 25 Jun 2020 18:18
 Operator : SY/MD
 Sample : L3121-11
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9F40

Quant Time: Jun 26 03:55:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 26 00:59:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	252844	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.43	117	253092	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	108426	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	79256	4.28	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.60%
7) Chloroethane-d5	1.93	69	76470	5.00	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.00%
11) 1,1-Dichloroethene-d2	2.59	63	121844	3.42	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.40%
20) 2-Butanone-d5	4.68	46	382851	57.29	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.58%
24) Chloroform-d	5.10	84	183568	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.80%
26) 1,2-Dichloroethane-d4	5.74	65	114019	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.80%
32) Benzene-d6	5.76	84	339330	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.80%
36) 1,2-Dichloropropane-d6	6.72	67	114259	4.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.80%
41) Toluene-d8	7.92	98	254404	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.60%
43) trans-1,3-Dichloropropene-	8.20	79	45229	4.41	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.20%
46) 2-Hexanone-d5	8.66	63	275213	52.02	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.04%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	102253	4.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.20%
64) 1,2-Dichlorobenzene-d4	12.20	152	106625	5.28	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.60%

Target Compounds					Ovalue
25) Chloroform	5.12	83	69109	1.847 ug/L	92
38) Bromodichloromethane	7.13	83	12355	0.460 ug/L	88
47) Tetrachloroethene	8.57	164	1510	0.109 ug/L	90
49) Dibromochloromethane	8.83	129	2307	0.132 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062520\
Data File : VU039124.D
Acq On : 25 Jun 2020 18:18
Operator : SY/MD
Sample : L3121-11
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9F40

Quant Time: Jun 26 03:55:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR062320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 26 00:59:08 2020
Response via : Initial Calibration

